

gas mixtures HgCl_2 solutions are used in liquid evaporative Hg gas generators. The results obtained were incorporated in the calibration protocol developed for Hg^0 gas generators.

To determine if the developed calibration protocols are fit for purpose validation measurements were performed by evaluating Hg gas generators available on the market. During the performance evaluation candidate generators were calibrated according to the protocols and based on the measurement data characteristics were determined, e.g., the stabilisation period, short-term drift, precision, i.e., reproducibility and repeatability of the concentration generated, linearity, bias, sensitivity to sample gas pressure, sensitivity to surrounding temperature and sensitivity to electrical voltage. All Hg gas generators could be tested according to the calibration protocols developed within the project. The results obtained with the different Hg gas generators clearly show the importance of a metrological calibration (Figure 2).

The SI-Hg project successfully developed calibration protocols for Hg^0 and Hg^{II} gas generators. Validation of the protocols was obtained through the performance evaluation of Hg gas generators available on the market. Based on the results calibration protocols, validation reports and good practice guides were written. These documents are available online (www.SI-Hg.eu and/or Search Metrology for traceable protocols for elemental and oxidised mercury concentrations (zenodo.org)). The documents have been handed over to CEN/TC 264 "Air Quality" WG8 "Measurement of Total Mercury Emissions". This standardisation committee will work on the conversion of the protocols into written documentary standards. Once, the documentary standards are in place they will replace other non-comparable references used in industry to calibrate Hg gas generators and ensure SI traceable measurement results for Hg concentrations in emission sources and the atmosphere. This is essential to underpin global efforts to control and reduce the concentration of Hg in the environment, comply with legislation and protect human health.

This project (19NRM03 SI-Hg) has received funding from the EMPIR programme co-financed by the Participating States and from the European Union's Horizon 2020 research and innovation programme.

Acknowledgments go to the SI-Hg project consortium, VSL, JSI, LGC Limited, VTT, CNR, Lumex Analytical GmbH, P. S. Analytical Limited and TÜV Rheinland Energy GmbH.




Figure 2: Results for the relative deviation (D %) of the Hg^0 concentration from the candidate generator setpoint against the calibrated output of the candidate generator traceable to the VSL primary standard. Generator 1 and Generator 2 are automatic saturation gas generators from different suppliers. The bell-jar was tested in the range from 50 ng m^{-3} to 300 ng m^{-3} with a deviation of +7%. Generator 1 was tested in the range from 3000 ng m^{-3} till 20000 ng m^{-3} with a deviation of -5% and Generator 2 was tested in the range from 5000 ng m^{-3} till 25000 ng m^{-3} with a deviation of -0.5%. Furthermore, the setpoints of Generator 2 are traceable to NIST. These results show comparability between the VSL primary Hg gas standard and a NIST calibrated gas generator.

Author Contact Details

: Iris de Krom – Scientist • VSL

• Address: Thijsseweg 11 Delft, The Netherlands • Tel: 0031 (0)15 269 1500

• Email: idekrom@vsl.nl • Web: www.vsl.nl/en

